

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032123\
 Data File : VW025465.D
 Acq On : 21 Mar 2023 13:52
 Operator : SY/MD
 Sample : 01979-12
 Misc : 5.58g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EYGB7

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/22/2023
 Supervised By :Mahesh Dadoda 03/22/2023

Quant Time: Mar 22 01:33:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Mar 22 01:30:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	948736	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	821146	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	351143	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	186649	11.832	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	47.320%		
7) Chloroethane-d5	2.899	69	184985	16.297	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	65.200%		
11) 1,1-Dichloroethene-d2	4.027	63	404939	13.291	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	53.160%		
21) 2-Butanone-d5	7.075	46	283143	73.125	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	146.260%#		
24) Chloroform-d	7.648	84	615234	22.494	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	89.960%		
26) 1,2-Dichloroethane-d4	8.307	65	383370m	26.115	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	104.440%		
32) Benzene-d6	8.276	84	1255165	22.370	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	89.480%		
36) 1,2-Dichloropropane-d6	9.270	67	426780	24.587	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	98.360%		
41) Toluene-d8	10.325	98	1087688	21.987	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	87.960%		
43) trans-1,3-Dichloroprop...	10.575	79	176494	25.535	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	102.160%		
47) 2-Hexanone-d5	10.922	63	192356	77.066	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	154.140%#		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	387102	30.440	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	121.760%#		
66) 1,2-Dichlorobenzene-d4	13.848	152	335304	25.678	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	102.720%		
Target Compounds					Qvalue	
13) Acetone	4.119	43	86417	23.217	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032123\
Data File : VW025465.D
Acq On : 21 Mar 2023 13:52
Operator : SY/MD
Sample : 01979-12
Misc : 5.58g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EYGB7

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/22/2023
Supervised By : Mahesh Dadoda 03/22/2023

Quant Time: Mar 22 01:33:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Mar 22 01:30:00 2023
Response via : Initial Calibration

