

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032123\
 Data File : VW025463.D
 Acq On : 21 Mar 2023 13:04
 Operator : SY/MD
 Sample : VIBLK424
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK424

Quant Time: Mar 22 01:33:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Mar 22 01:30:00 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	781871	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	683777	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	320778	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.362	65	236309	18.177	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.720%
7) Chloroethane-d5	2.899	69	188196	20.119	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.480%
11) 1,1-Dichloroethene-d2	4.021	63	395203	15.739	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.960%
21) 2-Butanone-d5	7.075	46	155280	48.662	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.320%
24) Chloroform-d	7.648	84	507174	22.501	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.000%
26) 1,2-Dichloroethane-d4	8.306	65	281383	23.258	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.040%
32) Benzene-d6	8.276	84	1059912	22.686	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.760%
36) 1,2-Dichloropropane-d6	9.270	67	335414	23.206	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.840%
41) Toluene-d8	10.324	98	931420	22.611	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	90.440%
43) trans-1,3-Dichloroprop...	10.574	79	126125	21.914	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.640%
47) 2-Hexanone-d5	10.922	63	99380	47.815	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.620%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	242736	22.923	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.680%
66) 1,2-Dichlorobenzene-d4	13.848	152	286127	23.986	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	95.960%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032123\
Data File : VW025463.D
Acq On : 21 Mar 2023 13:04
Operator : SY/MD
Sample : VIBLK424
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK424

Quant Time: Mar 22 01:33:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Mar 22 01:30:00 2023
Response via : Initial Calibration

