

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032124\
 Data File : VW028223.D
 Acq On : 21 Mar 2024 15:58
 Operator : SY/MD
 Sample : P1812-08
 Misc : 5.02g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Quant Time: Mar 22 01:32:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Mar 22 01:31:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	211134	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	188898	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	97673	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	48661	19.698	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	78.800%
7) Chloroethane-d5	2.905	69	36267	21.013	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	84.040%
11) 1,1-Dichloroethene-d2	4.027	65	32577	20.795	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	83.200%
21) 2-Butanone-d5	7.087	46	33959	46.608	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.220%
24) Chloroform-d	7.648	84	142029	22.813	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	91.240%
26) 1,2-Dichloroethane-d4	8.307	65	80288	23.401	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.600%
32) Benzene-d6	8.276	84	282580	24.502	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.000%
36) 1,2-Dichloropropane-d6	9.276	67	81446	24.540	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	98.160%
41) Toluene-d8	10.325	98	257142	24.173	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	96.680%
43) trans-1,3-Dichloroprop...	10.575	79	35593	23.612	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	94.440%
47) 2-Hexanone-d5	10.922	63	26255	47.858	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.720%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	59145	24.125	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.480%
66) 1,2-Dichlorobenzene-d4	13.848	152	85430	24.010	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	96.040%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032124\
Data File : VW028223.D
Acq On : 21 Mar 2024 15:58
Operator : SY/MD
Sample : P1812-08
Misc : 5.02g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: Mar 22 01:32:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Mar 22 01:31:33 2024
Response via : Initial Calibration

