

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032524\
 Data File : VW028239.D
 Acq On : 25 Mar 2024 10:27
 Operator : SY/MD
 Sample : VW0325SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK451

Quant Time: Mar 26 01:18:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Mar 26 01:17:39 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	210835	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	190193	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	95684	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.368	65	44973	18.231	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.920%
7) Chloroethane-d5	2.905	69	35497	20.596	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.400%
11) 1,1-Dichloroethene-d2	4.033	65	29878	19.099	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	76.400%
21) 2-Butanone-d5	7.081	46	30802	42.335	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.660%
24) Chloroform-d	7.654	84	133052	21.401	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	85.600%
26) 1,2-Dichloroethane-d4	8.306	65	73174	21.358	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.440%
32) Benzene-d6	8.276	84	262733	22.626	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	90.520%
36) 1,2-Dichloropropane-d6	9.276	67	75964	22.732	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	90.920%
41) Toluene-d8	10.324	98	238224	22.242	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.960%
43) trans-1,3-Dichloroprop...	10.580	79	31804	20.955	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	83.800%
47) 2-Hexanone-d5	10.922	63	24699	44.715	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	89.440%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	54058	21.899	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.600%
66) 1,2-Dichlorobenzene-d4	13.848	152	78993	22.663	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.640%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032524\
Data File : VW028239.D
Acq On : 25 Mar 2024 10:27
Operator : SY/MD
Sample : VW0325SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK451

Quant Time: Mar 26 01:18:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Mar 26 01:17:39 2024
Response via : Initial Calibration

