

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032724\
 Data File : VW028250.D
 Acq On : 27 Mar 2024 15:01
 Operator : SY/MD
 Sample : P1886-06
 Misc : 5.38g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E07A5

Quant Time: Mar 28 01:13:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 28 01:11:29 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	404531	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	357831	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	164968	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	63316	13.377	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.520%
7) Chloroethane-d5	2.899	69	49583	14.994	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.960%
11) 1,1-Dichloroethene-d2	4.021	65	44332	14.770	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	59.080%
21) 2-Butanone-d5	7.075	46	51581	36.949	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	73.900%
24) Chloroform-d	7.648	84	206724	17.330	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	69.320%
26) 1,2-Dichloroethane-d4	8.307	65	111412	16.948	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	67.800%#
32) Benzene-d6	8.276	84	399880	18.304	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	73.200%
36) 1,2-Dichloropropane-d6	9.270	67	117553	18.698	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	74.800%
41) Toluene-d8	10.325	98	357807	17.756	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	71.040%
43) trans-1,3-Dichloroprop...	10.575	79	49653	17.389	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	69.560%
47) 2-Hexanone-d5	10.922	63	40730	39.193	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.380%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	88456	19.047	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	76.200%
66) 1,2-Dichlorobenzene-d4	13.849	152	115322	19.190	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	76.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032724\
Data File : VW028250.D
Acq On : 27 Mar 2024 15:01
Operator : SY/MD
Sample : P1886-06
Misc : 5.38g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E07A5

Quant Time: Mar 28 01:13:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 28 01:11:29 2024
Response via : Initial Calibration

