

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032824\
 Data File : VW028262.D
 Acq On : 28 Mar 2024 10:12
 Operator : SY/MD
 Sample : VW0328SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK453

Quant Time: Mar 28 23:32:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 28 23:31:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	215328	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	195686	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	97372	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	47349	18.793	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.160%
7) Chloroethane-d5	2.905	69	38482	21.862	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	87.440%
11) 1,1-Dichloroethene-d2	4.027	65	32242	20.181	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	80.720%
21) 2-Butanone-d5	7.081	46	38345	51.603	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.200%
24) Chloroform-d	7.654	84	151241	23.819	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	95.280%
26) 1,2-Dichloroethane-d4	8.307	65	82779	23.657	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.640%
32) Benzene-d6	8.276	84	296771	24.840	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	99.360%
36) 1,2-Dichloropropane-d6	9.276	67	89691	26.087	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.360%
41) Toluene-d8	10.325	98	268188	24.337	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	97.360%
43) trans-1,3-Dichloroprop...	10.575	79	35970	23.034	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	92.120%
47) 2-Hexanone-d5	10.922	63	29571	52.033	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	104.060%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	64764	25.500	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	102.000%
66) 1,2-Dichlorobenzene-d4	13.849	152	95476	26.917	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	107.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032824\
Data File : VW028262.D
Acq On : 28 Mar 2024 10:12
Operator : SY/MD
Sample : VW0328SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK453

Quant Time: Mar 28 23:32:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 28 23:31:04 2024
Response via : Initial Calibration

