

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032824\
 Data File : VW028270.D
 Acq On : 28 Mar 2024 15:05
 Operator : SY/MD
 Sample : P1886-20
 Misc : 3.86g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E07B6

Quant Time: Mar 28 23:34:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 28 23:31:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	449762	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	392735	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	177083	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	71359	13.560	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	54.240%
7) Chloroethane-d5	2.899	69	54915	14.936	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.760%
11) 1,1-Dichloroethene-d2	4.021	65	46209	13.847	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	55.400%
21) 2-Butanone-d5	7.075	46	58434	37.648	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.300%
24) Chloroform-d	7.648	84	210952	15.906	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	63.640%
26) 1,2-Dichloroethane-d4	8.307	65	114002	15.598	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	62.400%#
32) Benzene-d6	8.276	84	417095	17.395	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	69.560%
36) 1,2-Dichloropropane-d6	9.270	67	128935	18.685	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	74.760%
41) Toluene-d8	10.319	98	373147	16.872	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	67.480%
43) trans-1,3-Dichloroprop...	10.575	79	51262	16.357	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	65.440%
47) 2-Hexanone-d5	10.922	63	47527	41.669	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	83.340%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	94072	18.456	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	73.840%
66) 1,2-Dichlorobenzene-d4	13.849	152	117268	18.179	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	72.720%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032824\
Data File : VW028270.D
Acq On : 28 Mar 2024 15:05
Operator : SY/MD
Sample : P1886-20
Misc : 3.86g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E07B6

Quant Time: Mar 28 23:34:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 28 23:31:04 2024
Response via : Initial Calibration

