

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032923\
 Data File : VW025494.D
 Acq On : 29 Mar 2023 18:07
 Operator : SY/MD
 Sample : 02121-10
 Misc : 5.42g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Mar 30 00:24:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Mar 30 00:21:45 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	995520	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	887125	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	404892	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	159772	9.652	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	38.600%
7) Chloroethane-d5	2.893	69	167654	14.076	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	56.320%
11) 1,1-Dichloroethene-d2	4.021	63	367957	11.509	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	46.040%
21) 2-Butanone-d5	7.075	46	187493	46.147	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	92.300%
24) Chloroform-d	7.648	84	585111	20.388	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	81.560%
26) 1,2-Dichloroethane-d4	8.410	65	345	0.022	ug/L	0.10
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.080%#
32) Benzene-d6	8.276	84	1186000	19.566	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	78.280%
36) 1,2-Dichloropropane-d6	9.270	67	395524	21.092	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	84.360%
41) Toluene-d8	10.319	98	1048005	19.609	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	78.440%
43) trans-1,3-Dichloroprop...	10.575	79	149461	20.016	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	80.080%
47) 2-Hexanone-d5	10.922	63	122824	45.549	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.100%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	295421	21.503	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	86.000%
66) 1,2-Dichlorobenzene-d4	13.848	152	338198	22.461	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032923\
Data File : VW025494.D
Acq On : 29 Mar 2023 18:07
Operator : SY/MD
Sample : 02121-10
Misc : 5.42g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Mar 30 00:24:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM030623SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Mar 30 00:21:45 2023
Response via : Initial Calibration

