

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032924\
 Data File : VW028275.D
 Acq On : 29 Mar 2024 10:26
 Operator : SY/MD
 Sample : VW0329SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK454

Quant Time: Mar 30 01:03:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Mar 30 00:57:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	229165	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	213100	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	107337	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.369	65	50472	18.823	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.280%
7) Chloroethane-d5	2.905	69	41207	21.997	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.000%
11) 1,1-Dichloroethene-d2	4.027	65	32239	18.960	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	75.840%
21) 2-Butanone-d5	7.081	46	40821	51.618	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.240%
24) Chloroform-d	7.654	84	157298	23.278	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.120%
26) 1,2-Dichloroethane-d4	8.307	65	85551	22.973	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.880%
32) Benzene-d6	8.276	84	309139	23.761	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.040%
36) 1,2-Dichloropropane-d6	9.276	67	97674	26.087	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.360%
41) Toluene-d8	10.325	98	273277	22.772	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.080%
43) trans-1,3-Dichloroprop...	10.575	79	37251	21.905	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	87.640%
47) 2-Hexanone-d5	10.922	63	30370	49.072	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.140%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	71820	25.968	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	103.880%
66) 1,2-Dichlorobenzene-d4	13.848	152	100360	25.667	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	102.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032924\
Data File : VW028275.D
Acq On : 29 Mar 2024 10:26
Operator : SY/MD
Sample : VW0329SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK454

Quant Time: Mar 30 01:03:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Mar 30 00:57:23 2024
Response via : Initial Calibration

