

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032924\
 Data File : VW028278.D
 Acq On : 29 Mar 2024 14:41
 Operator : SY/MD
 Sample : P1933-03
 Misc : 5.22g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EZZZ5

Quant Time: Mar 30 01:04:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Mar 30 00:57:23 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	325721	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	260390	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	86813	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	56142	14.731	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.920%
7) Chloroethane-d5	2.905	69	43157	16.209	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.840%
11) 1,1-Dichloroethene-d2	4.021	65	34841	14.416	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.680%
21) 2-Butanone-d5	7.081	46	65056	57.877	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	115.760%
24) Chloroform-d	7.648	84	166982	17.385	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	69.560%
26) 1,2-Dichloroethane-d4	8.307	65	103254	19.508	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	78.040%
32) Benzene-d6	8.276	84	326650	20.547	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	82.200%
36) 1,2-Dichloropropane-d6	9.270	67	104343	22.807	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.240%
41) Toluene-d8	10.319	98	265146	18.082	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.320%
43) trans-1,3-Dichloroprop...	10.575	79	42847	20.620	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.480%
47) 2-Hexanone-d5	10.922	63	51857	68.573	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	137.140%#
56) 1,1,2,2-Tetrachloroeth...	12.690	84	92053	27.238	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	108.960%
66) 1,2-Dichlorobenzene-d4	13.849	152	62139	19.649	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW032924\
Data File : VW028278.D
Acq On : 29 Mar 2024 14:41
Operator : SY/MD
Sample : P1933-03
Misc : 5.22g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EZZZ5

Quant Time: Mar 30 01:04:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Mar 30 00:57:23 2024
Response via : Initial Calibration

