

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW033119\
 Data File : VW009610.D
 Acq On : 31 Mar 2019 15:00
 Operator : SY/VA
 Sample : K2130-02
 Misc : 6.38G/5.0ML/MSVOA W/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FK-GF-02-034-A

Quant Time: Apr 01 05:24:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W032919S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 29 13:24:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	82769	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	124256	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	107149	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	42770	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	49432	52.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.46%	
35) Dibromofluoromethane	7.88	113	37037	47.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.92%	
50) Toluene-d8	10.32	98	119664	41.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.96%	
62) 4-Bromofluorobenzene	12.62	95	34982	32.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.02%	
Target Compounds						
16) Acetone	4.12	43	20176	94.810	ug/l	96
17) Carbon Disulfide	4.35	76	4926	2.167	ug/l #	81
18) Methyl Acetate	4.68	43	3193	6.207	ug/l	94
20) Methylene Chloride	4.90	84	4243	1.305	ug/l #	56
25) 2-Butanone	7.15	43	9034	30.720	ug/l #	66
80) 1,3,5-Trimethylbenzene	12.94	105	5882	2.347	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW033119\
Data File : VW009610.D
Acq On : 31 Mar 2019 15:00
Operator : SY/VA
Sample : K2130-02
Misc : 6.38G/5.0ML/MSVOA W/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FK-GF-02-034-A

Quant Time: Apr 01 05:24:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W032919S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 29 13:24:53 2019
Response via : Initial Calibration

