

Data Path : Z:\VOASRV\HPCHEM1\MSVOA W\DATA\VW033119\
 Data File : VW009611.D
 Acq On : 31 Mar 2019 15:27
 Operator : SY/VA
 Sample : K2130-04
 Misc : 6.92G/5.0ML/MSVOA W/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 FK-GF-02-034-B

Manual Integrations
 APPROVED

MMDadoda
 4/1/2019 5:30:30 PM

Quant Time: Apr 01 05:28:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_W\METHOD\82W032919S.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 29 13:24:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.95	168	252284	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.84	114	416358	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.63	117	381016	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.56	152	168759	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.30	65	145343	50.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.76%	
35) Dibromofluoromethane	7.88	113	117277	44.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.70%	
50) Toluene-d8	10.32	98	401931	41.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.16%	
62) 4-Bromofluorobenzene	12.62	95	129291	35.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	71.72%	
Target Compounds						
16) Acetone	4.13	43	18725	28.868	ug/l #	79
17) Carbon Disulfide	4.37	76	7625	1.100	ug/l #	93
20) Methylene Chloride	4.92	84	13284m	1.454	ug/l	
52) Toluene	10.38	92	14821	2.029	ug/l	98
80) 1,3,5-Trimethylbenzene	12.94	105	15630	1.580	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA W\DATA\VW033119\
Data File : VW009611.D
Acq On : 31 Mar 2019 15:27
Operator : SY/VA
Sample : K2130-04
Misc : 6.92G/5.0ML/MSVOA W/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
FK-GF-02-034-B

Manual Integrations
APPROVED

MMDadoda
4/1/2019 5:30:30 PM

Quant Time: Apr 01 05:28:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_W\METHOD\82W032919S.M
Quant Title : SW846 8260
QLast Update : Fri Mar 29 13:24:53 2019
Response via : Initial Calibration

