

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040122\
 Data File : VW022367.D
 Acq On : 01 Apr 2022 15:31
 Operator : SY/VA
 Sample : VW0401SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK471

Quant Time: Apr 02 02:35:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Apr 02 02:06:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	264702	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.628	117	254659	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	131654	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	97768	22.537	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.160%
7) Chloroethane-d5	2.891	69	63230	24.415	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.640%
11) 1,1-Dichloroethene-d2	4.025	63	118799	18.456	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.840%
21) 2-Butanone-d5	7.080	46	54227	51.156	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.320%
24) Chloroform-d	7.653	84	180836	24.501	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	98.000%
26) 1,2-Dichloroethane-d4	8.305	65	115862	26.548	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	106.200%
32) Benzene-d6	8.275	84	370920	26.524	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	106.080%
36) 1,2-Dichloropropane-d6	9.274	67	110550	26.332	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.320%
41) Toluene-d8	10.323	98	350649	26.356	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	105.440%
43) trans-1,3-Dichloroprop...	10.579	79	53864	25.513	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	102.040%
47) 2-Hexanone-d5	10.920	63	45913	53.387	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	106.780%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	89250	25.199	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.800%
66) 1,2-Dichlorobenzene-d4	13.853	152	113702	26.753	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	107.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040122\
Data File : VW022367.D
Acq On : 01 Apr 2022 15:31
Operator : SY/VA
Sample : VW0401SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK471

Quant Time: Apr 02 02:35:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040122SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Apr 02 02:06:11 2022
Response via : Initial Calibration

