

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040323\
 Data File : VW025552.D
 Acq On : 03 Apr 2023 13:07
 Operator : SY/MD
 Sample : 02151-08
 Misc : 5.89g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C11B1

Quant Time: Apr 04 03:11:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040323SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Apr 04 03:04:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	954786	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	765770	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	269512	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	319255	20.548	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.200%
7) Chloroethane-d5	2.899	69	228505	20.657	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	82.640%
11) 1,1-Dichloroethene-d2	4.021	63	494658	15.743	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	62.960%
21) 2-Butanone-d5	7.075	46	161354	46.552	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.100%
24) Chloroform-d	7.648	84	550700	20.542	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	82.160%
26) 1,2-Dichloroethane-d4	8.307	65	299882	21.502	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	86.000%
32) Benzene-d6	8.276	84	1174042	23.678	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	94.720%
36) 1,2-Dichloropropane-d6	9.270	67	362937	23.475	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	93.880%
41) Toluene-d8	10.325	98	994253	22.955	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.800%
43) trans-1,3-Dichloroprop...	10.575	79	125166	21.594	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	86.360%
47) 2-Hexanone-d5	10.922	63	104385	49.396	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	98.800%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	229447	21.988	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.960%
66) 1,2-Dichlorobenzene-d4	13.849	152	207247	22.383	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.520%
Target Compounds						
					Qvalue	
13) Acetone	4.119	43	107290	36.574	ug/L	97
16) Methylene chloride	4.917	84	50454	3.324	ug/L	95
22) 2-Butanone	7.173	43	33308	7.987	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040323\
Data File : VW025552.D
Acq On : 03 Apr 2023 13:07
Operator : SY/MD
Sample : 02151-08
Misc : 5.89g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C11B1

Quant Time: Apr 04 03:11:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040323SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Apr 04 03:04:42 2023
Response via : Initial Calibration

