

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040324\
 Data File : VW028298.D
 Acq On : 03 Apr 2024 17:09
 Operator : SY/MD
 Sample : P1965-07
 Misc : 4.34g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 E07D7

Quant Time: Apr 04 03:05:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 04 03:01:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	408894	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	368200	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.555	152	179569	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.368	65	71646	14.975	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	59.920%
7) Chloroethane-d5	2.905	69	56247	16.828	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.320%
11) 1,1-Dichloroethene-d2	4.021	65	46326	15.270	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.080%
21) 2-Butanone-d5	7.081	46	53399	37.843	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.680%
24) Chloroform-d	7.648	84	216907	17.990	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	71.960%
26) 1,2-Dichloroethane-d4	8.441	65	72	0.011	ug/L	0.14
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.040%#
32) Benzene-d6	8.276	84	420654	18.712	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	74.840%
36) 1,2-Dichloropropane-d6	9.276	67	129232	19.976	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	79.920%
41) Toluene-d8	10.324	98	376540	18.160	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	72.640%
43) trans-1,3-Dichloroprop...	10.574	79	50181	17.079	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.320%
47) 2-Hexanone-d5	10.922	63	41699	38.995	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.000%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	93228	19.509	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.040%
66) 1,2-Dichlorobenzene-d4	13.848	152	131059	20.035	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	80.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040324\
Data File : VW028298.D
Acq On : 03 Apr 2024 17:09
Operator : SY/MD
Sample : P1965-07
Misc : 4.34g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
E07D7

Quant Time: Apr 04 03:05:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Apr 04 03:01:50 2024
Response via : Initial Calibration

