

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040324\
 Data File : VW028294.D
 Acq On : 03 Apr 2024 15:33
 Operator : SY/MD
 Sample : VIBLK422
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK422

Quant Time: Apr 04 03:04:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 04 03:01:50 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	224484	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	209123	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	104361	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	43070	16.398	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	65.600%
7) Chloroethane-d5	2.899	69	37054	20.192	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	80.760%
11) 1,1-Dichloroethene-d2	4.021	65	28148	16.899	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.600%
21) 2-Butanone-d5	7.075	46	47327	61.092	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	122.180%
24) Chloroform-d	7.654	84	147762	22.322	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.280%
26) 1,2-Dichloroethane-d4	8.307	65	83925	23.007	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	92.040%
32) Benzene-d6	8.276	84	279777	21.913	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	87.640%
36) 1,2-Dichloropropane-d6	9.276	67	90980	24.761	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.040%
41) Toluene-d8	10.325	98	245763	20.869	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	83.480%
43) trans-1,3-Dichloroprop...	10.575	79	37181	22.280	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.120%
47) 2-Hexanone-d5	10.922	63	36589	60.245	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	120.480%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	78144	28.791	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	115.160%
66) 1,2-Dichlorobenzene-d4	13.849	152	96380	25.352	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	101.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040324\
Data File : VW028294.D
Acq On : 03 Apr 2024 15:33
Operator : SY/MD
Sample : VIBLK422
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK422

Quant Time: Apr 04 03:04:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM031424SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Apr 04 03:01:50 2024
Response via : Initial Calibration

