

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040422\
 Data File : VW022390.D
 Acq On : 04 Apr 2022 16:40
 Operator : SY/VA
 Sample : N2179-04
 Misc : 4.43g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0R43

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/05/2022
 Supervised By :Mahesh Dadoda 04/06/2022

Quant Time: Apr 05 01:57:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040122SMA.M
 Quant Title : SFAM01.0
 QLast Update : Tue Apr 05 01:44:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	208439	25.000	ug/L	# 0.00
28) Chlorobenzene-d5	11.633	117	190379	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	76469	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	103329	30.248	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	121.000%		
7) Chloroethane-d5	2.885	69	74717	36.637	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	146.560%		
11) 1,1-Dichloroethene-d2	4.019	63	92138	18.178	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	72.720%		
21) 2-Butanone-d5	7.073	46	49403	59.185	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	118.360%		
24) Chloroform-d	7.646	84	149523	25.727	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	102.920%		
26) 1,2-Dichloroethane-d4	8.305	65	100573m	29.265	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	117.080%		
32) Benzene-d6	8.274	84	311203	29.768	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	119.080%		
36) 1,2-Dichloropropane-d6	9.274	67	93470	29.781	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	119.120%		
41) Toluene-d8	10.323	98	293936	29.553	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	118.200%		
43) trans-1,3-Dichloroprop...	10.579	79	40917	25.925	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	103.680%		
47) 2-Hexanone-d5	10.920	63	43478	67.625	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	135.260%#		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	80895	30.552	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	122.200%#		
66) 1,2-Dichlorobenzene-d4	13.853	152	71083	28.795	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	115.200%		
Target Compounds					Qvalue	
13) Acetone	4.123	43	22472	39.315	ug/L	96
22) 2-Butanone	7.171	43	9587	10.633	ug/L	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW040422\
Data File : VW022390.D
Acq On : 04 Apr 2022 16:40
Operator : SY/VA
Sample : N2179-04
Misc : 4.43g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0R43

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 04/05/2022
Supervised By : Mahesh Dadoda 04/06/2022

Quant Time: Apr 05 01:57:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040122SMA.M
Quant Title : SFAM01.0
QLast Update : Tue Apr 05 01:44:25 2022
Response via : Initial Calibration

