

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041124\
 Data File : VW028410.D
 Acq On : 11 Apr 2024 17:48
 Operator : SY/MD
 Sample : P2078-05
 Misc : 4.08g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AA9

Quant Time: Apr 12 01:00:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 12 00:55:02 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	197344	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	175673	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	69380	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	76396	24.147	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.600%
7) Chloroethane-d5	2.905	69	60630	25.285	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.120%
11) 1,1-Dichloroethene-d2	4.021	65	32007	23.890	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	95.560%
21) 2-Butanone-d5	7.081	46	30393	51.338	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.680%
24) Chloroform-d	7.648	84	127740	24.806	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	99.240%
26) 1,2-Dichloroethane-d4	8.410	65	82	0.031	ug/L	0.10
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.120%#
32) Benzene-d6	8.276	84	253929	26.459	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.840%
36) 1,2-Dichloropropane-d6	9.276	67	77260	27.084	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.320%
41) Toluene-d8	10.325	98	219477	25.150	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.600%
43) trans-1,3-Dichloroprop...	10.575	79	27007	23.810	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	95.240%
47) 2-Hexanone-d5	10.922	63	20258	51.268	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	102.540%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	55993	27.821	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.280%
66) 1,2-Dichlorobenzene-d4	13.849	152	62101	28.303	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	113.200%
Target Compounds						
13) Acetone	4.131	43	9362	12.879	ug/L	Qvalue 87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041124\
Data File : VW028410.D
Acq On : 11 Apr 2024 17:48
Operator : SY/MD
Sample : P2078-05
Misc : 4.08g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AA9

Quant Time: Apr 12 01:00:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 12 00:55:02 2024
Response via : Initial Calibration





#13
Acetone
Concen: 12.879 ug/L
RT: 4.131 min Scan# 30
Delta R.T. 0.006 min
Lab File: VW028410.D
Acq: 11 Apr 2024 17:48

Instrument :
MSVOA_W
ClientSampleId :
C0AA9

Tgt Ion:	43	Resp:	9362
Ion	Ratio	Lower	Upper
43	100		
58	33.6	0.0	53.4

