

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041124\
 Data File : VW028414.D
 Acq On : 11 Apr 2024 19:23
 Operator : SY/MD
 Sample : P2078-12
 Misc : 5.26g/10mL/MSVOA_W/SOIL/A
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 C0AC9

Quant Time: Apr 12 01:01:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 12 00:55:02 2024
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.843	114	272796	25.000	ug/L	0.00
28) Chlorobenzene-d5		11.629	117	245135	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.556	152	94259	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.350	65	127766	29.215	ug/L	-0.02
Spiked Amount	25.000	Range	30 - 150	Recovery	= 116.840%		
7) Chloroethane-d5		2.887	69	100010	30.172	ug/L	-0.02
Spiked Amount	25.000	Range	30 - 150	Recovery	= 120.680%		
11) 1,1-Dichloroethene-d2		4.003	65	46371	25.038	ug/L	-0.02
Spiked Amount	25.000	Range	45 - 110	Recovery	= 100.160%		
21) 2-Butanone-d5		7.100	46	51106	62.448	ug/L	0.02
Spiked Amount	50.000	Range	20 - 135	Recovery	= 124.900%		
24) Chloroform-d		7.648	84	196129	27.552	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 110.200%		
26) 1,2-Dichloroethane-d4		8.307	65	112443	30.396	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 121.600%		
32) Benzene-d6		8.270	84	380238	28.394	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 113.560%		
36) 1,2-Dichloropropane-d6		9.270	67	118485	29.766	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 119.080%		
41) Toluene-d8		10.325	98	328450	26.973	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 107.880%		
43) trans-1,3-Dichloroprop...		10.575	79	40800	25.778	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 103.120%		
47) 2-Hexanone-d5		10.922	63	33426	60.622	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 121.240%		
56) 1,1,2,2-Tetrachloroeth...		12.690	84	86517	30.806	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 123.240%#		
66) 1,2-Dichlorobenzene-d4		13.848	152	85771	28.773	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 115.080%		
Target Compounds							Qvalue
13) Acetone		4.173	43	8631	8.589	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041124\
Data File : VW028414.D
Acq On : 11 Apr 2024 19:23
Operator : SY/MD
Sample : P2078-12
Misc : 5.26g/10mL/MSVOA_W/SOIL/A
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
C0AC9

Quant Time: Apr 12 01:01:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 12 00:55:02 2024
Response via : Initial Calibration

