

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041224\
 Data File : VW028432.D
 Acq On : 12 Apr 2024 15:39
 Operator : SY/MD
 Sample : P2078-17RE
 Misc : 4.67g/10mL/MSVOA_W/SOIL/B
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: Apr 12 23:15:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040824SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 12 23:09:31 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	435115	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	342929	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.550	152	124635	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.351	65	152372	21.844	ug/L	-0.02
Spiked Amount	25.000	Range 30 - 150	Recovery	=	87.360%	
7) Chloroethane-d5	2.887	69	119152	22.537	ug/L	-0.01
Spiked Amount	25.000	Range 30 - 150	Recovery	=	90.160%	
11) 1,1-Dichloroethene-d2	3.997	65	59400	20.108	ug/L	-0.02
Spiked Amount	25.000	Range 45 - 110	Recovery	=	80.440%	
21) 2-Butanone-d5	7.106	46	66508	50.952	ug/L	0.02
Spiked Amount	50.000	Range 20 - 135	Recovery	=	101.900%	
24) Chloroform-d	7.648	84	251747	22.172	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	88.680%	
26) 1,2-Dichloroethane-d4	8.051	65	99	0.017	ug/L	-0.25
Spiked Amount	25.000	Range 70 - 130	Recovery	=	0.080%#	
32) Benzene-d6	8.270	84	505579	26.987	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	107.960%	
36) 1,2-Dichloropropane-d6	9.270	67	155743	27.968	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	111.880%	
41) Toluene-d8	10.319	98	432899	25.412	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	101.640%	
43) trans-1,3-Dichloroprop...	10.575	79	49495	22.353	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	89.400%	
47) 2-Hexanone-d5	10.928	63	48417	62.769	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	125.540%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	113007	28.763	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	115.040%	
66) 1,2-Dichlorobenzene-d4	13.849	152	112341	28.501	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	114.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041224\
Data File : VW028432.D
Acq On : 12 Apr 2024 15:39
Operator : SY/MD
Sample : P2078-17RE
Misc : 4.67g/10mL/MSVOA_W/SOIL/B
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: Apr 12 23:15:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040824SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 12 23:09:31 2024
Response via : Initial Calibration

