

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041322\
 Data File : VW022528.D
 Acq On : 13 Apr 2022 13:41
 Operator : SY/VA
 Sample : N2374-07
 Misc : 5.25g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFFE3

Quant Time: Apr 14 01:32:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 14 01:27:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	337638	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	281711	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	101600	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	120900	16.197	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	64.800%
7) Chloroethane-d5	2.892	69	89901	19.218	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.880%
11) 1,1-Dichloroethene-d2	4.019	63	131794	13.613	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.440%
21) 2-Butanone-d5	7.080	46	62870	48.245	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.480%
24) Chloroform-d	7.647	84	236190	22.380	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.520%
26) 1,2-Dichloroethane-d4	8.305	65	142709	23.667	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.680%
32) Benzene-d6	8.275	84	416404	24.193	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.760%
36) 1,2-Dichloropropane-d6	9.274	67	131509	26.403	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.600%
41) Toluene-d8	10.323	98	363932	21.590	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.360%
43) trans-1,3-Dichloroprop...	10.579	79	48719	20.726	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.920%
47) 2-Hexanone-d5	10.921	63	51900	56.006	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	112.020%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	111914	26.425	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.720%
66) 1,2-Dichlorobenzene-d4	13.853	152	80375	22.335	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041322\
Data File : VW022528.D
Acq On : 13 Apr 2022 13:41
Operator : SY/VA
Sample : N2374-07
Misc : 5.25g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFFE3

Quant Time: Apr 14 01:32:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Apr 14 01:27:44 2022
Response via : Initial Calibration

