

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041322\
 Data File : VW022532.D
 Acq On : 13 Apr 2022 15:24
 Operator : SY/VA
 Sample : N2374-08
 Misc : 4.23g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFFE4

Quant Time: Apr 14 01:33:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu Apr 14 01:27:44 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	269632	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	202368	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	68247	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	100326	16.831	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	67.320%
7) Chloroethane-d5	2.885	69	84056	22.501	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.000%
11) 1,1-Dichloroethene-d2	4.019	63	102596	13.270	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	53.080%
21) 2-Butanone-d5	7.080	46	54960	52.812	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	105.620%
24) Chloroform-d	7.647	84	188029	22.311	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	89.240%
26) 1,2-Dichloroethane-d4	8.305	65	115631	24.013	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.040%
32) Benzene-d6	8.275	84	326621	26.417	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.680%
36) 1,2-Dichloropropane-d6	9.274	67	102865	28.750	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	115.000%
41) Toluene-d8	10.323	98	260956	21.550	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	10.579	79	37661	22.303	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.200%
47) 2-Hexanone-d5	10.927	63	44533	66.897	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	133.800%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	89296	29.352	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	117.400%
66) 1,2-Dichlorobenzene-d4	13.853	152	54731	22.641	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	90.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041322\
Data File : VW022532.D
Acq On : 13 Apr 2022 15:24
Operator : SY/VA
Sample : N2374-08
Misc : 4.23g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFFE4

Quant Time: Apr 14 01:33:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Thu Apr 14 01:27:44 2022
Response via : Initial Calibration

