

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041422\
 Data File : VW022546.D
 Acq On : 14 Apr 2022 13:44
 Operator : SY/VA
 Sample : N2373-10
 Misc : 4.76g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFFE7

Quant Time: Apr 15 01:31:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 15 01:29:34 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	295485	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	240433	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	81658	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	94972	14.539	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	58.160%		
7) Chloroethane-d5	2.891	69	69130	16.886	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	67.560%		
11) 1,1-Dichloroethene-d2	4.013	63	92740	10.946	ug/L	-0.01
Spiked Amount	25.000	Range 45 - 110	Recovery =	43.800%#		
21) 2-Butanone-d5	7.074	46	70678	61.974	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	123.940%		
24) Chloroform-d	7.647	84	190125	20.586	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	82.360%		
26) 1,2-Dichloroethane-d4	8.305	65	114539	21.705	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	86.840%		
32) Benzene-d6	8.275	84	327227	22.276	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	89.120%		
36) 1,2-Dichloropropane-d6	9.274	67	107330	25.248	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	101.000%		
41) Toluene-d8	10.323	98	271357	18.862	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	75.440%		
43) trans-1,3-Dichloroprop...	10.573	79	31385	15.644	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	62.560%		
47) 2-Hexanone-d5	10.921	63	53338	67.439	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	134.880%		
56) 1,1,2,2-Tetrachloroeth...	12.695	84	103575	28.655	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	114.640%		
66) 1,2-Dichlorobenzene-d4	13.853	152	68272	23.604	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	94.400%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041422\
 Data File : VW022546.D
 Acq On : 14 Apr 2022 13:44
 Operator : SY/VA
 Sample : N2373-10
 Misc : 4.76g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFFE7

Quant Time: Apr 15 01:31:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 15 01:29:34 2022
 Response via : Initial Calibration

