

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041522\
 Data File : VW022557.D
 Acq On : 15 Apr 2022 14:38
 Operator : SY/VA
 Sample : N2373-11RE
 Misc : 5.58g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 BFFFORE

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/17/2022
 Supervised By :Mahesh Dadoda 04/20/2022

Quant Time: Apr 15 23:18:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 15 23:17:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	257390	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	215962	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	80035	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	84402	14.833	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	59.320%		
7) Chloroethane-d5	2.885	69	63406	17.780	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	71.120%		
11) 1,1-Dichloroethene-d2	4.013	63	84625	11.466	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	45.880%		
21) 2-Butanone-d5	7.080	46	53163	53.515	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	107.040%		
24) Chloroform-d	7.647	84	160459	19.945	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	79.760%		
26) 1,2-Dichloroethane-d4	8.305	65	99505m	21.647	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	86.600%		
32) Benzene-d6	8.275	84	275317	20.866	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	83.480%		
36) 1,2-Dichloropropane-d6	9.274	67	89217	23.365	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	93.480%		
41) Toluene-d8	10.323	98	223845	17.322	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	69.280%		
43) trans-1,3-Dichloroprop...	10.579	79	32465	18.016	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	72.080%		
47) 2-Hexanone-d5	10.921	63	39924	56.199	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	112.400%		
56) 1,1,2,2-Tetrachloroeth...	12.689	84	82905	25.535	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	102.160%		
66) 1,2-Dichlorobenzene-d4	13.853	152	56758	20.022	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	80.080%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW041522\
Data File : VW022557.D
Acq On : 15 Apr 2022 14:38
Operator : SY/VA
Sample : N2373-11RE
Misc : 5.58g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
BFFF0RE

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/17/2022
Supervised By :Mahesh Dadoda 04/20/2022

Quant Time: Apr 15 23:18:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 15 23:17:23 2022
Response via : Initial Calibration

