

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042023\
 Data File : VW025603.D
 Acq On : 20 Apr 2023 14:05
 Operator : SY/MD
 Sample : 02418-05
 Misc : 5.32g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW8Z1

Quant Time: Apr 21 01:50:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040323SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 21 01:47:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	676395	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	546296	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	188699	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	179835	16.338	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	65.360%	
7) Chloroethane-d5	2.899	69	159084	20.300	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	81.200%	
11) 1,1-Dichloroethene-d2	4.021	63	314290	14.119	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	56.480%	
21) 2-Butanone-d5	7.075	46	200126	81.501	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	163.000%#	
24) Chloroform-d	7.648	84	483945	25.482	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	101.920%	
26) 1,2-Dichloroethane-d4	8.307	65	299798	30.343	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	121.360%	
32) Benzene-d6	8.276	84	942992	26.658	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	106.640%	
36) 1,2-Dichloropropane-d6	9.276	67	331389	30.046	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	120.200%#	
41) Toluene-d8	10.325	98	762921	24.690	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	98.760%	
43) trans-1,3-Dichloroprop...	10.575	79	120896	29.236	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	116.960%	
47) 2-Hexanone-d5	10.922	63	126574	83.958	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	167.920%#	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	280065	37.621	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	150.480%#	
66) 1,2-Dichlorobenzene-d4	13.848	152	185141	28.559	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	114.240%	
Target Compounds					Qvalue	
13) Acetone	4.118	43	16909	8.136	ug/L	91
16) Methylene chloride	4.929	84	22218	2.067	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042023\
Data File : VW025603.D
Acq On : 20 Apr 2023 14:05
Operator : SY/MD
Sample : 02418-05
Misc : 5.32g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW8Z1

Quant Time: Apr 21 01:50:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040323SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 21 01:47:12 2023
Response via : Initial Calibration

