

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042523\
 Data File : VW025674.D
 Acq On : 25 Apr 2023 11:27
 Operator : SY/MD
 Sample : 02479-10
 Misc : 5.23g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW951

Quant Time: Apr 26 01:14:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040323SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Apr 26 01:13:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	796946	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	580429	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	160267	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	173529	13.381	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	53.520%
7) Chloroethane-d5	2.899	69	164908	17.860	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	71.440%
11) 1,1-Dichloroethene-d2	4.027	63	318516	12.145	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	48.560%
21) 2-Butanone-d5	7.075	46	157418	54.411	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	108.820%
24) Chloroform-d	7.649	84	493752	22.066	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	88.280%
26) 1,2-Dichloroethane-d4	8.307	65	277460	23.834	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	95.320%
32) Benzene-d6	8.270	84	940111	25.014	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	100.040%
36) 1,2-Dichloropropane-d6	9.270	67	321456	27.431	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	109.720%
41) Toluene-d8	10.325	98	729844	22.231	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.920%
43) trans-1,3-Dichloroprop...	10.575	79	99732	22.700	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.800%
47) 2-Hexanone-d5	10.922	63	94130	58.766	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.540%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	208302	26.336	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	105.360%
66) 1,2-Dichlorobenzene-d4	13.849	152	128575	23.352	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.400%
Target Compounds					Qvalue	
13) Acetone	4.125	43	15459	6.313	ug/L	93
16) Methylene chloride	4.923	84	48160	3.802	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042523\
Data File : VW025674.D
Acq On : 25 Apr 2023 11:27
Operator : SY/MD
Sample : 02479-10
Misc : 5.23g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW951

Quant Time: Apr 26 01:14:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040323SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Apr 26 01:13:04 2023
Response via : Initial Calibration

