

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042523\
 Data File : VW025687.D
 Acq On : 25 Apr 2023 16:58
 Operator : SY/MD
 Sample : 02479-07RE
 Misc : 5.18g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW948RE

Quant Time: Apr 26 01:17:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040323SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Apr 26 01:13:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.837	114	803571	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	532215	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	110169	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	165013	12.619	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	50.480%
7) Chloroethane-d5	2.899	69	159056	17.084	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.320%
11) 1,1-Dichloroethene-d2	4.021	63	303144	11.463	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	45.840%
21) 2-Butanone-d5	7.075	46	138565	47.500	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.000%
24) Chloroform-d	7.648	84	491712	21.793	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.160%
26) 1,2-Dichloroethane-d4	8.307	65	266250	22.683	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.720%
32) Benzene-d6	8.276	84	911490	26.449	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	105.800%
36) 1,2-Dichloropropane-d6	9.276	67	317122	29.513	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	118.040%
41) Toluene-d8	10.325	98	674578	22.409	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	89.640%
43) trans-1,3-Dichloroprop...	10.575	79	89183	22.138	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.560%
47) 2-Hexanone-d5	10.922	63	83659	56.960	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	113.920%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	181925	25.084	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.320%
66) 1,2-Dichlorobenzene-d4	13.848	152	84550	22.339	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	89.360%
Target Compounds						
					Qvalue	
13) Acetone	4.125	43	8169	3.309	ug/L	90
16) Methylene chloride	4.911	84	64266	5.031	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042523\
Data File : VW025687.D
Acq On : 25 Apr 2023 16:58
Operator : SY/MD
Sample : 02479-07RE
Misc : 5.18g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW948RE

Quant Time: Apr 26 01:17:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040323SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Apr 26 01:13:04 2023
Response via : Initial Calibration

