

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042622\
 Data File : VW022616.D
 Acq On : 26 Apr 2022 10:52
 Operator : SY/VA
 Sample : VW0426SBL01
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VBLK486

Quant Time: Apr 27 01:38:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Apr 27 01:34:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	286777	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	280743	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	149596	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	137318	21.660	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	86.640%
7) Chloroethane-d5	2.885	69	92712	23.334	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	93.320%
11) 1,1-Dichloroethene-d2	4.013	63	151128	18.379	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	73.520%
21) 2-Butanone-d5	7.080	46	55972	50.569	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	101.140%
24) Chloroform-d	7.653	84	225713	25.181	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.720%
26) 1,2-Dichloroethane-d4	8.525	65	64	0.012	ug/L	0.22
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.040%#
32) Benzene-d6	8.275	84	442595	25.804	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.200%
36) 1,2-Dichloropropane-d6	9.274	67	125774	25.339	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.360%
41) Toluene-d8	10.323	98	423974	25.238	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.960%
43) trans-1,3-Dichloroprop...	10.579	79	52073	22.229	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.920%
47) 2-Hexanone-d5	10.921	63	44970	48.695	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	97.380%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	106050	25.127	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	100.520%
66) 1,2-Dichlorobenzene-d4	13.853	152	140640	26.542	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042622\
Data File : VW022616.D
Acq On : 26 Apr 2022 10:52
Operator : SY/VA
Sample : VW0426SBL01
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VBLK486

Quant Time: Apr 27 01:38:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Apr 27 01:34:11 2022
Response via : Initial Calibration

