

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042723\
 Data File : VW025735.D
 Acq On : 27 Apr 2023 19:36
 Operator : SY/MD
 Sample : 02499-11
 Misc : 4.96g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCCG3

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 04/28/2023
 Supervised By :Mahesh Dadoda 05/01/2023

Quant Time: Apr 28 03:04:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM042723SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 28 02:55:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	650016	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	583617	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	268958	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	119374	15.055	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	60.200%		
7) Chloroethane-d5	2.893	69	98831	17.385	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	69.520%		
11) 1,1-Dichloroethene-d2	4.027	63	222210m	12.292	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	49.160%		
21) 2-Butanone-d5	7.075	46	155198	59.314	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	118.620%		
24) Chloroform-d	7.648	84	326324	18.714	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	74.840%		
26) 1,2-Dichloroethane-d4	8.307	65	199757m	21.013	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	84.040%		
32) Benzene-d6	8.270	84	625308	17.648	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	70.600%		
36) 1,2-Dichloropropane-d6	9.270	67	214773	18.740	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	74.960%		
41) Toluene-d8	10.319	98	543533	17.352	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	69.400%		
43) trans-1,3-Dichloroprop...	10.575	79	82589	18.492	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	73.960%		
47) 2-Hexanone-d5	10.922	63	102188	57.408	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	114.820%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	203085	24.043	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	96.160%		
66) 1,2-Dichlorobenzene-d4	13.849	152	185253	20.104	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	80.400%		
Target Compounds					Qvalue	
13) Acetone	4.119	43	135222	55.754	ug/L	99
16) Methylene chloride	4.923	84	26671	2.172	ug/L	96
22) 2-Butanone	7.179	43	8380	2.725	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042723\
Data File : VW025735.D
Acq On : 27 Apr 2023 19:36
Operator : SY/MD
Sample : 02499-11
Misc : 4.96g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCCG3

Quant Time: Apr 28 03:04:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM042723SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 28 02:55:08 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/28/2023
Supervised By :Mahesh Dadoda 05/01/2023

