

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042823\
 Data File : VW025760.D
 Acq On : 28 Apr 2023 15:17
 Operator : SY/MD
 Sample : 02526-15
 Misc : 5.99g/10mL/MSVOA_W/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCCF9

Quant Time: Apr 29 00:24:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM042723SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 28 02:55:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	602611	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	551552	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	265983	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	130268	17.721	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	70.880%
7) Chloroethane-d5	2.808	69	271	0.051	ug/L	-0.09
Spiked Amount	25.000	Range	30 - 150	Recovery	=	0.200%#
11) 1,1-Dichloroethene-d2	4.027	63	244951	14.616	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	58.480%
21) 2-Butanone-d5	7.075	46	159752	65.857	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	131.720%
24) Chloroform-d	7.648	84	366817	22.691	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.760%
26) 1,2-Dichloroethane-d4	8.307	65	227076	25.766	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	103.080%
32) Benzene-d6	8.276	84	696436	20.798	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	83.200%
36) 1,2-Dichloropropane-d6	9.270	67	239996	22.158	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	88.640%
41) Toluene-d8	10.319	98	596145	20.138	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	80.560%
43) trans-1,3-Dichloroprop...	10.575	79	94570	22.405	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.640%
47) 2-Hexanone-d5	10.922	63	107658	63.997	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	128.000%
56) 1,1,2,2-Tetrachloroeth...	12.690	84	221625	27.764	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	111.040%
66) 1,2-Dichlorobenzene-d4	13.848	152	221492	24.306	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	97.240%
Target Compounds						
13) Acetone	4.118	43	54664	24.312	ug/L	Qvalue 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042823\
Data File : VW025760.D
Acq On : 28 Apr 2023 15:17
Operator : SY/MD
Sample : 02526-15
Misc : 5.99g/10mL/MSVOA_W/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCCF9


Quant Time: Apr 29 00:24:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM042723SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 28 02:55:08 2023
Response via : Initial Calibration





#13
 Acetone
 Concen: 24.312 ug/L
 RT: 4.118 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: VW025760.D
 Acq: 28 Apr 2023 15:17

Instrument :
 MSVOA_W
 ClientSampleId :
 DCCF9



Tgt Ion: 43 Resp: 54664
 Ion Ratio Lower Upper
 43 100
 58 30.7 0.0 58.6

