

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042922\
 Data File : VW022689.D
 Acq On : 29 Apr 2022 10:45
 Operator : SY/VA
 Sample : N2603-12
 Misc : 7.02g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW493

Quant Time: Apr 29 22:23:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 29 22:19:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	228381	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	203427	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	78561	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	54116	10.719	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	42.880%
7) Chloroethane-d5	2.885	69	49608	15.678	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	62.720%
11) 1,1-Dichloroethene-d2	4.013	63	67242	10.268	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	41.080%#
21) 2-Butanone-d5	7.080	46	60542	68.684	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	137.360%#
24) Chloroform-d	7.647	84	141128	19.770	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.080%
26) 1,2-Dichloroethane-d4	8.305	65	85059	20.855	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	83.400%
32) Benzene-d6	8.275	84	238011	19.150	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	76.600%
36) 1,2-Dichloropropane-d6	9.274	67	80211	22.301	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	89.200%
41) Toluene-d8	10.323	98	197766	16.247	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	65.000%
43) trans-1,3-Dichloroprop...	10.579	79	28548	16.818	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	67.280%
47) 2-Hexanone-d5	10.920	63	46605	69.645	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	139.300%#
56) 1,1,2,2-Tetrachloroeth...	12.688	84	91583	29.947	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	119.800%
66) 1,2-Dichlorobenzene-d4	13.847	152	65078	23.387	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	93.560%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042922\
Data File : VW022689.D
Acq On : 29 Apr 2022 10:45
Operator : SY/VA
Sample : N2603-12
Misc : 7.02g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW493

Quant Time: Apr 29 22:23:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 29 22:19:25 2022
Response via : Initial Calibration

