

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042922\
 Data File : VW022725.D
 Acq On : 30 Apr 2022 02:54
 Operator : SY/VA
 Sample : N2638-20
 Misc : 5.58g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXL13

Quant Time: Apr 30 04:03:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Apr 30 01:41:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	273936	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	273524	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	131150	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	200678	33.138	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 132.560%			
7) Chloroethane-d5	2.885	69	144325	38.027	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery = 152.120%#			
11) 1,1-Dichloroethene-d2	4.019	63	215976	27.497	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery = 110.000%#			
21) 2-Butanone-d5	7.080	46	106262	100.505	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 201.020%#			
24) Chloroform-d	7.647	84	382707	44.697	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery = 178.800%#			
26) 1,2-Dichloroethane-d4	8.305	65	220172	45.005	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery = 180.000%#			
32) Benzene-d6	8.275	84	731903	43.797	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery = 175.200%#			
36) 1,2-Dichloropropane-d6	9.275	67	222974	46.107	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery = 184.440%#			
41) Toluene-d8	10.323	98	677788	41.412	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery = 165.640%#			
43) trans-1,3-Dichloroprop...	10.573	79	88588	38.815	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery = 155.240%#			
47) 2-Hexanone-d5	10.921	63	91414	101.598	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery = 203.200%#			
56) 1,1,2,2-Tetrachloroeth...	12.689	84	201116	48.909	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery = 195.640%#			
66) 1,2-Dichlorobenzene-d4	13.853	152	214108	46.091	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery = 184.360%#			
Target Compounds						Qvalue
46) Tetrachloroethene	10.860	164	38461	10.523	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042922\
Data File : VW022725.D
Acq On : 30 Apr 2022 02:54
Operator : SY/VA
Sample : N2638-20
Misc : 5.58g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXL13

Quant Time: Apr 30 04:03:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Apr 30 01:41:13 2022
Response via : Initial Calibration

