

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042922\
 Data File : VW022691.D
 Acq On : 29 Apr 2022 11:33
 Operator : SY/VA
 Sample : N2603-10RE
 Misc : 7.50g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW491RE

Quant Time: Apr 29 22:23:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Apr 29 22:19:25 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	147794	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	124249	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	43847	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	59947	18.348	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	73.400%
7) Chloroethane-d5	2.885	69	49213	24.034	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	96.120%
11) 1,1-Dichloroethene-d2	4.019	63	69665	16.439	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	65.760%
21) 2-Butanone-d5	7.080	46	52427	91.909	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	183.820%#
24) Chloroform-d	7.647	84	151945	32.892	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	131.560%
26) 1,2-Dichloroethane-d4	8.305	65	90210	34.178	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	136.720%#
32) Benzene-d6	8.275	84	254007	33.461	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	133.840%
36) 1,2-Dichloropropane-d6	9.275	67	85312	38.835	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	155.320%#
41) Toluene-d8	10.323	98	200306	26.942	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	107.760%
43) trans-1,3-Dichloroprop...	10.579	79	28544	27.532	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	110.120%
47) 2-Hexanone-d5	10.921	63	43197	105.689	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	211.380%#
56) 1,1,2,2-Tetrachloroeth...	12.689	84	88247	47.244	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	188.960%#
66) 1,2-Dichlorobenzene-d4	13.847	152	53651	34.545	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	138.200%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW042922\
Data File : VW022691.D
Acq On : 29 Apr 2022 11:33
Operator : SY/VA
Sample : N2603-10RE
Misc : 7.50g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW491RE

Quant Time: Apr 29 22:23:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM040722SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Apr 29 22:19:25 2022
Response via : Initial Calibration

