

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050322\
 Data File : VW022768.D
 Acq On : 03 May 2022 13:37
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: May 04 01:25:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 04 01:21:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	298672	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	291555	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	147565	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	121161	19.773	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	79.080%
7) Chloroethane-d5	2.885	69	91713	23.006	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	92.040%
11) 1,1-Dichloroethene-d2	4.019	63	118949	15.023	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.080%
21) 2-Butanone-d5	7.080	46	55510	47.645	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	95.300%
24) Chloroform-d	7.647	84	190855	21.811	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	87.240%
26) 1,2-Dichloroethane-d4	8.128	65	129	0.027	ug/L	-0.18
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.120%#
32) Benzene-d6	8.275	84	371761	22.122	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	88.480%
36) 1,2-Dichloropropane-d6	9.274	67	107974	22.873	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	91.480%
41) Toluene-d8	10.323	98	352092	21.689	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	86.760%
43) trans-1,3-Dichloroprop...	10.579	79	44655	20.545	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	82.160%
47) 2-Hexanone-d5	10.920	63	42515	47.179	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.360%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	102969	23.403	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.600%
66) 1,2-Dichlorobenzene-d4	13.853	152	121205	23.497	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050322\
Data File : VW022768.D
Acq On : 03 May 2022 13:37
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: May 04 01:25:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 04 01:21:13 2022
Response via : Initial Calibration

