

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050322\
 Data File : VW022766.D
 Acq On : 03 May 2022 12:49
 Operator : SY/VA
 Sample : VIBLK443
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK443

Quant Time: May 04 01:25:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 04 01:21:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	295400	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	285790	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	143640	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	129861	21.427	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.720%
7) Chloroethane-d5	2.885	69	87835	22.277	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	89.120%
11) 1,1-Dichloroethene-d2	4.019	63	123793	15.808	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.240%
21) 2-Butanone-d5	7.080	46	62880	54.569	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	109.140%
24) Chloroform-d	7.647	84	203913	23.561	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.240%
26) 1,2-Dichloroethane-d4	8.305	65	118920	25.053	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	100.200%
32) Benzene-d6	8.275	84	400117	24.290	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.160%
36) 1,2-Dichloropropane-d6	9.274	67	117044	25.295	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.160%
41) Toluene-d8	10.323	98	369043	23.192	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	92.760%
43) trans-1,3-Dichloroprop...	10.579	79	48137	22.594	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	90.360%
47) 2-Hexanone-d5	10.920	63	51771	58.609	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	117.220%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	111514	25.856	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	103.440%
66) 1,2-Dichlorobenzene-d4	13.853	152	126379	25.169	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	100.680%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050322\
Data File : VW022766.D
Acq On : 03 May 2022 12:49
Operator : SY/VA
Sample : VIBLK443
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK443

Quant Time: May 04 01:25:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 04 01:21:13 2022
Response via : Initial Calibration

