

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050322\
 Data File : VW022769.D
 Acq On : 03 May 2022 14:01
 Operator : SY/VA
 Sample : N2639-11
 Misc : 4.86g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXL21

Quant Time: May 04 01:26:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 04 01:21:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	336630	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	331825	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	163567	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	128451	18.599	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.400%
7) Chloroethane-d5	2.885	69	91294	20.318	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.280%
11) 1,1-Dichloroethene-d2	4.019	63	126620	14.189	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	56.760%
21) 2-Butanone-d5	7.080	46	61404	46.761	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	93.520%
24) Chloroform-d	7.653	84	194954	19.767	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	79.080%
26) 1,2-Dichloroethane-d4	8.305	65	115969	21.439	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	85.760%
32) Benzene-d6	8.275	84	380061	19.871	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	79.480%
36) 1,2-Dichloropropane-d6	9.274	67	112105	20.866	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	83.480%
41) Toluene-d8	10.323	98	357134	19.330	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	77.320%
43) trans-1,3-Dichloroprop...	10.579	79	46527	18.808	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.240%
47) 2-Hexanone-d5	10.927	63	46882	45.711	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	91.420%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	111027	22.172	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	88.680%
66) 1,2-Dichlorobenzene-d4	13.853	152	122863	21.488	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	85.960%
Target Compounds						Qvalue
46) Tetrachloroethene	10.859	164	5184	1.283	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050322\
Data File : VW022769.D
Acq On : 03 May 2022 14:01
Operator : SY/VA
Sample : N2639-11
Misc : 4.86g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXL21

Quant Time: May 04 01:26:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 04 01:21:13 2022
Response via : Initial Calibration

