

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050322\
 Data File : VW022779.D
 Acq On : 03 May 2022 18:02
 Operator : SY/VA
 Sample : N2632-09
 Misc : 7.58g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW4B1

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/04/2022
 Supervised By :Mahesh Dadoda 05/04/2022

Quant Time: May 04 01:29:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 04 01:21:13 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	248518	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	196581	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	59013	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	129885	25.474	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.880%
7) Chloroethane-d5	2.879	69	89162	26.879	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	107.520%
11) 1,1-Dichloroethene-d2	4.013	63	110500m	16.773	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.080%
21) 2-Butanone-d5	7.074	46	34606	35.697	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	71.400%
24) Chloroform-d	7.647	84	182900	25.120	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.480%
26) 1,2-Dichloroethane-d4	8.305	65	108282	27.116	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	108.480%
32) Benzene-d6	8.275	84	348617	30.767	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	123.080%
36) 1,2-Dichloropropane-d6	9.275	67	102511	32.207	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	128.840%#
41) Toluene-d8	10.323	98	291455	26.628	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	106.520%
43) trans-1,3-Dichloroprop...	10.579	79	36097	24.631	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	98.520%
47) 2-Hexanone-d5	10.927	63	31494	51.833	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.660%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	89013	30.005	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	120.000%
66) 1,2-Dichlorobenzene-d4	13.853	152	54754	26.542	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	106.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050322\
Data File : VW022779.D
Acq On : 03 May 2022 18:02
Operator : SY/VA
Sample : N2632-09
Misc : 7.58g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW4B1

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/04/2022
Supervised By : Mahesh Dadoda 05/04/2022

Quant Time: May 04 01:29:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 04 01:21:13 2022
Response via : Initial Calibration

