

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050323\
 Data File : VW025825.D
 Acq On : 03 May 2023 14:33
 Operator : SY/MD
 Sample : 02574-07
 Misc : 5.93g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCCP2

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/04/2023
 Supervised By : Mahesh Dadoda 05/04/2023

Quant Time: May 04 01:26:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu May 04 01:22:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	948378	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	842058	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	375966	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	316625	23.560	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	94.240%		
7) Chloroethane-d5	2.899	69	231867	23.272	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery =	93.080%		
11) 1,1-Dichloroethene-d2	4.021	65	147647	21.940	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery =	87.760%		
21) 2-Butanone-d5	7.075	46	179065	40.304	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	80.600%		
24) Chloroform-d	7.648	84	584697	21.381	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery =	85.520%		
26) 1,2-Dichloroethane-d4	8.307	65	332970m	21.532	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery =	86.120%		
32) Benzene-d6	8.270	84	1229507	22.346	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery =	89.400%		
36) 1,2-Dichloropropane-d6	9.270	67	391984	22.266	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery =	89.080%		
41) Toluene-d8	10.319	98	1076373	21.938	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery =	87.760%		
43) trans-1,3-Dichloroprop...	10.575	79	147259	20.636	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery =	82.560%		
47) 2-Hexanone-d5	10.922	63	123377	41.100	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery =	82.200%		
56) 1,1,2,2-Tetrachloroeth...	12.690	84	283950	21.162	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery =	84.640%		
66) 1,2-Dichlorobenzene-d4	13.848	152	330546	24.118	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery =	96.480%		
Target Compounds					Qvalue	
13) Acetone	4.119	43	108055	27.429	ug/L	99
16) Methylene chloride	4.917	84	34206	1.749	ug/L	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050323\
Data File : VW025825.D
Acq On : 03 May 2023 14:33
Operator : SY/MD
Sample : 02574-07
Misc : 5.93g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCCP2

Quant Time: May 04 01:26:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050123SMA.M
Quant Title : SFAM01.0
QLast Update : Thu May 04 01:22:04 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/04/2023
Supervised By :Mahesh Dadoda 05/04/2023

