

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050323\
 Data File : VW025833.D
 Acq On : 03 May 2023 17:45
 Operator : SY/MD
 Sample : 02574-15
 Misc : 5.59g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 DCCP8

Quant Time: May 04 01:28:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu May 04 01:22:04 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.843	114	888100	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.629	117	798496	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.556	152	347871	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.363	65	285830	22.712	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	90.840%	
7) Chloroethane-d5	2.893	69	208883	22.388	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	89.560%	
11) 1,1-Dichloroethene-d2	4.021	65	134235	21.301	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	85.200%	
21) 2-Butanone-d5	7.075	46	184174	44.268	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	88.540%	
24) Chloroform-d	7.648	84	537825	21.002	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	84.000%	
26) 1,2-Dichloroethane-d4	8.307	65	318765	22.013	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	88.040%	
32) Benzene-d6	8.276	84	1111419	21.302	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	85.200%	
36) 1,2-Dichloropropane-d6	9.270	67	358423	21.471	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	85.880%	
41) Toluene-d8	10.325	98	989899	21.276	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	85.120%	
43) trans-1,3-Dichloroprop...	10.575	79	137227	20.279	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	81.120%	
47) 2-Hexanone-d5	10.922	63	127745	44.876	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	89.760%	
56) 1,1,2,2-Tetrachloroeth...	12.690	84	281577	22.130	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	88.520%	
66) 1,2-Dichlorobenzene-d4	13.848	152	299099	23.586	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	94.360%	
Target Compounds						
					Qvalue	
13) Acetone	4.119	43	175891	47.679	ug/L	100
22) 2-Butanone	7.179	43	13802	2.710	ug/L	84
46) Tetrachloroethene	10.861	164	14344	1.438	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050323\
Data File : VW025833.D
Acq On : 03 May 2023 17:45
Operator : SY/MD
Sample : 02574-15
Misc : 5.59g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
DCCP8

Quant Time: May 04 01:28:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050123SMA.M
Quant Title : SFAM01.0
QLast Update : Thu May 04 01:22:04 2023
Response via : Initial Calibration

