

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050422\
 Data File : VW022799.D
 Acq On : 04 May 2022 14:05
 Operator : SY/VA
 Sample : N2632-14RE
 Misc : 7.66g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW4B6RE

Quant Time: May 04 23:36:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 04 23:28:03 2022
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards							
1) 1,4-Difluorobenzene		8.842	114	273750	25.000	ug/L	0.00
28) Chlorobenzene-d5		11.628	117	174413	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		13.554	152	33126	25.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		2.355	65	140754	25.061	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 100.240%		
7) Chloroethane-d5		2.885	69	101584	27.801	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	= 111.200%		
11) 1,1-Dichloroethene-d2		4.013	63	124033	17.091	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	= 68.360%		
21) 2-Butanone-d5		7.074	46	55650	52.114	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 104.220%		
24) Chloroform-d		7.647	84	200225	24.965	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	= 99.840%		
26) 1,2-Dichloroethane-d4		8.305	65	109543	24.903	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	= 99.600%		
32) Benzene-d6		8.275	84	368445	36.650	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	= 146.600%#		
36) 1,2-Dichloropropane-d6		9.275	67	108366	38.374	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	= 153.480%#		
41) Toluene-d8		10.323	98	271133	27.919	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	= 111.680%		
43) trans-1,3-Dichloroprop...		10.579	79	30013	23.082	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	= 92.320%		
47) 2-Hexanone-d5		10.921	63	36962	68.565	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	= 137.120%#		
56) 1,1,2,2-Tetrachloroeth...		12.689	84	82905	31.498	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	= 126.000%#		
66) 1,2-Dichlorobenzene-d4		13.853	152	34901	30.140	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	= 120.560%#		
Target Compounds							Qvalue
13) Acetone		4.129	43	4572	5.719	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050422\
Data File : VW022799.D
Acq On : 04 May 2022 14:05
Operator : SY/VA
Sample : N2632-14RE
Misc : 7.66g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW4B6RE

Quant Time: May 04 23:36:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 04 23:28:03 2022
Response via : Initial Calibration

