

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050422\
 Data File : VW022790.D
 Acq On : 04 May 2022 10:28
 Operator : SY/VA
 Sample : N2632-01
 Misc : 7.80g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW4A3

Quant Time: May 04 23:35:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed May 04 23:28:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	199538	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	147568	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	36454	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	93128	22.748	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	91.000%
7) Chloroethane-d5	2.885	69	67923	25.503	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	102.000%
11) 1,1-Dichloroethene-d2	4.013	63	93944	17.760	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	71.040%
21) 2-Butanone-d5	7.080	46	36684	47.130	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	94.260%
24) Chloroform-d	7.647	84	132250	22.622	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	90.480%
26) 1,2-Dichloroethane-d4	8.305	65	75546	23.562	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	94.240%
32) Benzene-d6	8.275	84	248414	29.206	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	116.840%
36) 1,2-Dichloropropane-d6	9.275	67	75876	31.757	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	127.040%#
41) Toluene-d8	10.323	98	206020	25.074	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	100.280%
43) trans-1,3-Dichloroprop...	10.579	79	23395	21.266	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.080%
47) 2-Hexanone-d5	10.921	63	24588	53.908	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	107.820%
56) 1,1,2,2-Tetrachloroeth...	12.689	84	64782	29.090	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	116.360%
66) 1,2-Dichlorobenzene-d4	13.853	152	37227	29.213	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	116.840%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050422\
Data File : VW022790.D
Acq On : 04 May 2022 10:28
Operator : SY/VA
Sample : N2632-01
Misc : 7.80g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW4A3

Quant Time: May 04 23:35:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Wed May 04 23:28:03 2022
Response via : Initial Calibration

