

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050522\
 Data File : VW022855.D
 Acq On : 05 May 2022 21:34
 Operator : SY/VA
 Sample : N2708-12
 Misc : 5.41g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXL38

Quant Time: May 06 02:56:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 06 02:14:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	311179	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	259040	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	83552	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	122503	19.188	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	76.760%
7) Chloroethane-d5	2.885	69	91853	22.115	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	88.440%
11) 1,1-Dichloroethene-d2	4.013	63	112701	13.662	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	54.640%
21) 2-Butanone-d5	7.079	46	48140	39.659	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.320%
24) Chloroform-d	7.646	84	196891	21.596	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.400%
26) 1,2-Dichloroethane-d4	8.305	65	113597	22.718	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.880%
32) Benzene-d6	8.274	84	358778	24.029	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	96.120%
36) 1,2-Dichloropropane-d6	9.274	67	109627	26.138	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	104.560%
41) Toluene-d8	10.323	98	316140	21.919	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	87.680%
43) trans-1,3-Dichloroprop...	10.579	79	37652	19.497	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	78.000%
47) 2-Hexanone-d5	10.920	63	35147	43.898	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	87.800%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	91196	23.328	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.320%
66) 1,2-Dichlorobenzene-d4	13.853	152	64130	21.957	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	87.840%
Target Compounds						Qvalue
46) Tetrachloroethene	10.859	164	8163	2.588	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050522\
Data File : VW022855.D
Acq On : 05 May 2022 21:34
Operator : SY/VA
Sample : N2708-12
Misc : 5.41g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXL38

Quant Time: May 06 02:56:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 06 02:14:27 2022
Response via : Initial Calibration

