

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050522\
 Data File : VW022864.D
 Acq On : 06 May 2022 01:11
 Operator : SY/VA
 Sample : N2708-21
 Misc : 4.61g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EXL48

Quant Time: May 06 02:58:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 06 02:14:27 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	349141	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	322802	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	152642	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	145381	20.296	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	81.200%
7) Chloroethane-d5	2.885	69	105576	22.655	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.600%
11) 1,1-Dichloroethene-d2	4.019	63	132224	14.286	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	57.160%
21) 2-Butanone-d5	7.080	46	55015	40.395	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.780%
24) Chloroform-d	7.647	84	238875	23.353	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	93.400%
26) 1,2-Dichloroethane-d4	8.305	65	130664	23.290	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	93.160%
32) Benzene-d6	8.275	84	451571	24.270	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	97.080%
36) 1,2-Dichloropropane-d6	9.274	67	135045	25.838	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	103.360%
41) Toluene-d8	10.323	98	411907	22.917	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	91.680%
43) trans-1,3-Dichloroprop...	10.579	79	47843	19.881	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	79.520%
47) 2-Hexanone-d5	10.927	63	45055	45.157	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	90.320%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	110951	22.776	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	91.120%
66) 1,2-Dichlorobenzene-d4	13.847	152	126077	23.628	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.520%
Target Compounds						
16) Methylene chloride	4.922	84	20719	4.056	ug/L	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050522\
Data File : VW022864.D
Acq On : 06 May 2022 01:11
Operator : SY/VA
Sample : N2708-21
Misc : 4.61g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EXL48

Quant Time: May 06 02:58:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 06 02:14:27 2022
Response via : Initial Calibration

