

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050622\
 Data File : VW022910.D
 Acq On : 07 May 2022 00:09
 Operator : SY/VA
 Sample : N2714-01
 Misc : 4.10g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1




Instrument :
 MSVOA_W
 ClientSampleId :
 EW2W4

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2022
 Supervised By :Mahesh Dadoda 05/10/2022

Quant Time: May 07 03:07:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 07 01:54:32 2022
 Response via : Initial Calibration

05/07/2022
 Supervised By :Mahesh
 Dadoda



TIC: VW022910.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

8.482min (+ 0.177) 0.02 ug/L

response 68

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	51.10	45.59
51.00	92.30	97.06
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050622\
 Data File : VW022910.D
 Acq On : 07 May 2022 00:09
 Operator : SY/VA
 Sample : N2714-01
 Misc : 4.10g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1




Instrument :
 MSVOA_W
 ClientSampleId :
 EW2W4

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2022
 Supervised By :Mahesh Dadoda 05/10/2022

Quant Time: May 07 03:07:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 07 01:54:32 2022
 Response via : Initial Calibration

05/07/2022
 Supervised By :Mahesh
 Dadoda



TIC: VW022910.D\data.ms

(26) 1,2-Dichloroethane-d4 (S)

8.305min (+ 0.000) 26.54 ug/L m

response 96687

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	51.10	0.03#
51.00	92.30	0.07#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050622\
 Data File : VW022910.D
 Acq On : 07 May 2022 00:09
 Operator : SY/VA
 Sample : N2714-01
 Misc : 4.10g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW2W4

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2022
 Supervised By :Mahesh Dadoda 05/10/2022

Quant Time: May 07 03:07:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 07 01:54:32 2022
 Response via : Initial Calibration

05/07/2022
 Supervised By :Mahesh
 Dadoda

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----05/10/2022						
Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	226740	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	190947	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	62043	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	92260	19.833	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	79.320%	
7) Chloroethane-d5	2.885	69	69526	22.973	ug/L	0.00
Spiked Amount 25.000	Range 30	- 150	Recovery	=	91.880%	
11) 1,1-Dichloroethene-d2	4.019	63	86344	14.365	ug/L	0.00
Spiked Amount 25.000	Range 45	- 110	Recovery	=	57.440%	
21) 2-Butanone-d5	7.080	46	59832	67.647	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	135.300%#	
24) Chloroform-d	7.647	84	150982	22.728	ug/L	0.00
Spiked Amount 25.000	Range 40	- 150	Recovery	=	90.920%	
26) 1,2-Dichloroethane-d4	8.305	65	96687m	26.538	ug/L	0.00
Spiked Amount 25.000	Range 70	- 130	Recovery	=	106.160%	
32) Benzene-d6	8.274	84	289779	26.329	ug/L	0.00
Spiked Amount 25.000	Range 20	- 135	Recovery	=	105.320%	
36) 1,2-Dichloropropane-d6	9.274	67	86654	28.029	ug/L	0.00
Spiked Amount 25.000	Range 70	- 120	Recovery	=	112.120%	
41) Toluene-d8	10.323	98	238882	22.468	ug/L	0.00
Spiked Amount 25.000	Range 30	- 130	Recovery	=	89.880%	
43) trans-1,3-Dichloroprop...	10.579	79	26540	18.644	ug/L	0.00
Spiked Amount 25.000	Range 30	- 135	Recovery	=	74.560%	
47) 2-Hexanone-d5	10.926	63	44462	75.335	ug/L	0.00
Spiked Amount 50.000	Range 20	- 135	Recovery	=	150.680%#	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	84896	29.461	ug/L	0.00
Spiked Amount 25.000	Range 45	- 120	Recovery	=	117.840%	
66) 1,2-Dichlorobenzene-d4	13.853	152	51390	23.695	ug/L	0.00
Spiked Amount 25.000	Range 75	- 120	Recovery	=	94.760%	
Target Compounds						
					Qvalue	
13) Acetone	4.123	43	42013	63.445	ug/L	100
22) 2-Butanone	7.171	43	16765	18.103	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050622\
Data File : VW022910.D
Acq On : 07 May 2022 00:09
Operator : SY/VA
Sample : N2714-01
Misc : 4.10g/10mL/MSVOA_W/soil
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW2W4

Manual IntegrationsAPPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2022
Supervised By :Mahesh Dadoda 05/10/2022

Quant Time: May 07 03:07:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 07 01:54:32 2022
Response via : Initial Calibration

05/07/2022
Supervised By :Mahesh
Dadoda

05/10/2022

