

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050622\
 Data File : VW022914.D
 Acq On : 07 May 2022 01:45
 Operator : SY/VA
 Sample : N2714-08
 Misc : 6.15g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW2X9

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 05/07/2022
 Supervised By : Mahesh Dadoda 05/10/2022

Quant Time: May 07 03:08:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat May 07 01:54:32 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	20381	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	18146	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	6074	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	7417	17.738	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	70.960%		
7) Chloroethane-d5	2.891	69	4834	17.770	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	71.080%		
11) 1,1-Dichloroethene-d2	4.038	63	7366	13.633	ug/L	0.02
Spiked Amount 25.000	Range 45 - 110		Recovery =	54.520%		
21) 2-Butanone-d5	7.092	46	2742m	34.489	ug/L	0.01
Spiked Amount 50.000	Range 20 - 135		Recovery =	68.980%		
24) Chloroform-d	7.647	84	13029	21.820	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	87.280%		
26) 1,2-Dichloroethane-d4	8.311	65	8300	25.344	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	101.360%		
32) Benzene-d6	8.281	84	21585	20.637	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	82.560%		
36) 1,2-Dichloropropane-d6	9.274	67	7733	26.320	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	105.280%		
41) Toluene-d8	10.323	98	17532	17.352	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	69.400%		
43) trans-1,3-Dichloroprop...	10.573	79	1896	14.016	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	56.080%		
47) 2-Hexanone-d5	10.933	63	897m	15.993	ug/L	0.01
Spiked Amount 50.000	Range 20 - 135		Recovery =	31.980%		
56) 1,1,2,2-Tetrachloroeth...	12.695	84	6310	23.042	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	92.160%		
66) 1,2-Dichlorobenzene-d4	13.853	152	3960	18.651	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	74.600%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050622\
Data File : VW022914.D
Acq On : 07 May 2022 01:45
Operator : SY/VA
Sample : N2714-08
Misc : 6.15g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW2X9

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/07/2022
Supervised By :Mahesh Dadoda 05/10/2022

Quant Time: May 07 03:08:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Sat May 07 01:54:32 2022
Response via : Initial Calibration

