

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050922\
 Data File : VW022930.D
 Acq On : 09 May 2022 15:16
 Operator : SY/VA
 Sample : VIBLK
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK453

Quant Time: May 09 22:58:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon May 09 22:55:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	287438	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.634	117	296500	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	147103	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	110309	18.705	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	74.840%	
7) Chloroethane-d5	2.891	69	83983	21.890	ug/L	0.00
Spiked Amount	25.000	Range 30 - 150	Recovery	=	87.560%	
11) 1,1-Dichloroethene-d2	4.019	63	119098	15.630	ug/L	0.00
Spiked Amount	25.000	Range 45 - 110	Recovery	=	62.520%	
21) 2-Butanone-d5	7.080	46	56401	50.302	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	100.600%	
24) Chloroform-d	7.653	84	199962	23.745	ug/L	0.00
Spiked Amount	25.000	Range 40 - 150	Recovery	=	94.960%	
26) 1,2-Dichloroethane-d4	8.311	65	110357	23.893	ug/L	0.00
Spiked Amount	25.000	Range 70 - 130	Recovery	=	95.560%	
32) Benzene-d6	8.275	84	380558	22.268	ug/L	0.00
Spiked Amount	25.000	Range 20 - 135	Recovery	=	89.080%	
36) 1,2-Dichloropropane-d6	9.274	67	112346	23.402	ug/L	0.00
Spiked Amount	25.000	Range 70 - 120	Recovery	=	93.600%	
41) Toluene-d8	10.323	98	357956	21.682	ug/L	0.00
Spiked Amount	25.000	Range 30 - 130	Recovery	=	86.720%	
43) trans-1,3-Dichloroprop...	10.579	79	44294	20.039	ug/L	0.00
Spiked Amount	25.000	Range 30 - 135	Recovery	=	80.160%	
47) 2-Hexanone-d5	10.926	63	42791	46.693	ug/L	0.00
Spiked Amount	50.000	Range 20 - 135	Recovery	=	93.380%	
56) 1,1,2,2-Tetrachloroeth...	12.688	84	105814	23.648	ug/L	0.00
Spiked Amount	25.000	Range 45 - 120	Recovery	=	94.600%	
66) 1,2-Dichlorobenzene-d4	13.853	152	123403	23.998	ug/L	0.00
Spiked Amount	25.000	Range 75 - 120	Recovery	=	96.000%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050922\
Data File : VW022930.D
Acq On : 09 May 2022 15:16
Operator : SY/VA
Sample : VIBLK
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK453

Quant Time: May 09 22:58:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Mon May 09 22:55:00 2022
Response via : Initial Calibration

