

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050922\
 Data File : VW022926.D
 Acq On : 09 May 2022 13:40
 Operator : SY/VA
 Sample : VIBLK451
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VIBLK451

Quant Time: May 09 22:57:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon May 09 22:55:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	305567	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	292935	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	150557	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	116746	18.622	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.480%
7) Chloroethane-d5	2.891	69	92113	22.585	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	90.320%
11) 1,1-Dichloroethene-d2	4.019	63	127897	15.789	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	63.160%
21) 2-Butanone-d5	7.080	46	57659	48.373	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.740%
24) Chloroform-d	7.647	84	211357	23.609	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	94.440%
26) 1,2-Dichloroethane-d4	8.305	65	118157	24.064	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	96.240%
32) Benzene-d6	8.275	84	403329	23.887	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	95.560%
36) 1,2-Dichloropropane-d6	9.274	67	120192	25.341	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	101.360%
41) Toluene-d8	10.323	98	380680	23.339	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.360%
43) trans-1,3-Dichloroprop...	10.579	79	48651	22.278	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	89.120%
47) 2-Hexanone-d5	10.920	63	46664	51.539	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	103.080%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	107031	24.211	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	96.840%
66) 1,2-Dichlorobenzene-d4	13.847	152	130636	24.822	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	99.280%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW050922\
Data File : VW022926.D
Acq On : 09 May 2022 13:40
Operator : SY/VA
Sample : VIBLK451
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VIBLK451

Quant Time: May 09 22:57:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Mon May 09 22:55:00 2022
Response via : Initial Calibration

