

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051122\
 Data File : VW022968.D
 Acq On : 11 May 2022 10:06
 Operator : SY/VA
 Sample : N2708-23
 Misc : 5.00g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 VHBLK001

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/15/2022
 Supervised By :Mahesh Dadoda 05/16/2022

Quant Time: May 12 01:08:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu May 12 01:07:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	305829	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.627	117	300287	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	150783	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	115457	18.401	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	73.600%		
7) Chloroethane-d5	2.885	69	85617	20.974	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	83.880%		
11) 1,1-Dichloroethene-d2	4.019	63	121612	15.000	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	60.000%		
21) 2-Butanone-d5	7.079	46	58820	49.305	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	98.600%		
24) Chloroform-d	7.646	84	215965	24.103	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	96.400%		
26) 1,2-Dichloroethane-d4	8.305	65	119292m	24.275	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	97.080%		
32) Benzene-d6	8.274	84	406612	23.492	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	93.960%		
36) 1,2-Dichloropropane-d6	9.274	67	122923	25.283	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	101.120%		
41) Toluene-d8	10.323	98	378654	22.647	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	90.600%		
43) trans-1,3-Dichloroprop...	10.579	79	47964	21.425	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	85.720%		
47) 2-Hexanone-d5	10.920	63	43419	46.781	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	93.560%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	110451	24.373	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	97.480%		
66) 1,2-Dichlorobenzene-d4	13.853	152	131756	24.997	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	100.000%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051122\
Data File : VW022968.D
Acq On : 11 May 2022 10:06
Operator : SY/VA
Sample : N2708-23
Misc : 5.00g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
VHBLK001

Quant Time: May 12 01:08:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Thu May 12 01:07:03 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 05/15/2022
Supervised By : Mahesh Dadoda 05/16/2022

