

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051122\
 Data File : VW022985.D
 Acq On : 11 May 2022 17:38
 Operator : SY/VA
 Sample : N2766-07
 Misc : 5.90g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :
 EW4F6

Quant Time: May 12 01:12:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Thu May 12 01:07:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.842	114	278090	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	248484	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.560	152	92332	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.355	65	95498	16.738	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	66.960%		
7) Chloroethane-d5	2.885	69	75650	20.381	ug/L	0.00
Spiked Amount 25.000	Range 30 - 150		Recovery =	81.520%		
11) 1,1-Dichloroethene-d2	4.019	63	98912	13.417	ug/L	0.00
Spiked Amount 25.000	Range 45 - 110		Recovery =	53.680%		
21) 2-Butanone-d5	7.080	46	36559	33.702	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	67.400%		
24) Chloroform-d	7.647	84	182192	22.362	ug/L	0.00
Spiked Amount 25.000	Range 40 - 150		Recovery =	89.440%		
26) 1,2-Dichloroethane-d4	8.305	65	94955	21.250	ug/L	0.00
Spiked Amount 25.000	Range 70 - 130		Recovery =	85.000%		
32) Benzene-d6	8.275	84	328183	22.914	ug/L	0.00
Spiked Amount 25.000	Range 20 - 135		Recovery =	91.640%		
36) 1,2-Dichloropropane-d6	9.274	67	100298	24.930	ug/L	0.00
Spiked Amount 25.000	Range 70 - 120		Recovery =	99.720%		
41) Toluene-d8	10.323	98	289192	20.902	ug/L	0.00
Spiked Amount 25.000	Range 30 - 130		Recovery =	83.600%		
43) trans-1,3-Dichloroprop...	10.579	79	32188	17.376	ug/L	0.00
Spiked Amount 25.000	Range 30 - 135		Recovery =	69.520%		
47) 2-Hexanone-d5	10.921	63	29246	38.079	ug/L	0.00
Spiked Amount 50.000	Range 20 - 135		Recovery =	76.160%		
56) 1,1,2,2-Tetrachloroeth...	12.688	84	82404	21.975	ug/L	0.00
Spiked Amount 25.000	Range 45 - 120		Recovery =	87.880%		
66) 1,2-Dichlorobenzene-d4	13.847	152	71252	22.076	ug/L	0.00
Spiked Amount 25.000	Range 75 - 120		Recovery =	88.320%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051122\
Data File : VW022985.D
Acq On : 11 May 2022 17:38
Operator : SY/VA
Sample : N2766-07
Misc : 5.90g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :
EW4F6

Quant Time: May 12 01:12:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Thu May 12 01:07:03 2022
Response via : Initial Calibration

