

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051222\
 Data File : VW022999.D
 Acq On : 12 May 2022 14:16
 Operator : SY/VA
 Sample : N2766-13
 Misc : 6.56g/10mL/MSVOA_W/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_W
 ClientSampleId :

Quant Time: May 13 01:12:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri May 13 01:10:05 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.841	114	357584	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.628	117	342167	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.554	152	153837	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.349	65	100155	13.652	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	54.600%
7) Chloroethane-d5	2.885	69	82084	17.198	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	68.800%
11) 1,1-Dichloroethene-d2	4.013	63	106830	11.270	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	45.080%
21) 2-Butanone-d5	7.074	46	52510	37.645	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	75.280%
24) Chloroform-d	7.647	84	200784	19.165	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	76.680%
26) 1,2-Dichloroethane-d4	8.506	65	143	0.025	ug/L	0.20
Spiked Amount	25.000	Range	70 - 130	Recovery	=	0.080%#
32) Benzene-d6	8.275	84	366185	18.567	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	74.280%
36) 1,2-Dichloropropane-d6	9.268	67	112823	20.365	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	81.480%
41) Toluene-d8	10.323	98	336506	17.663	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	70.640%
43) trans-1,3-Dichloroprop...	10.579	79	42659	16.723	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	66.880%
47) 2-Hexanone-d5	10.920	63	41568	39.305	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.600%
56) 1,1,2,2-Tetrachloroeth...	12.688	84	100455	19.454	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	77.800%
66) 1,2-Dichlorobenzene-d4	13.853	152	105465	19.612	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	78.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_W\Data\VW051222\
Data File : VW022999.D
Acq On : 12 May 2022 14:16
Operator : SY/VA
Sample : N2766-13
Misc : 6.56g/10mL/MSVOA_W/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_W
ClientSampleId :

Quant Time: May 13 01:12:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_W\Method\SFAMWLM050222SMA.M
Quant Title : SFAM01.0
QLast Update : Fri May 13 01:10:05 2022
Response via : Initial Calibration

